

CALIFORNIA AIR RESOURCES BOARD
MONTHLY QUALITY CONTROL MAINTENANCE CHECK SHEET
API Model 701 Zero Air Module

Location: _____ **Month/Year:** _____
Station Number: _____ **Operator:** _____
Property Number: _____ **Agency:** _____

OPERATOR INSTRUCTIONS:

1. Daily checks:

Check	ON	OFF
Power light (Red).		
Hygrometer light (Green).		

2. Monthly Checks:

Check	Recorded	Target
Check and record output pressure.		35 PSI \pm 1 PSI
Adjusted pressure (if applicable).		
Confirm the presence of water in the moisture output trap.		

3. Annual Checks:

Check	Last Recorded	Current
Change the activated carbon filter as needed or at a minimum of once per year.		
Change the Purafil® filter as needed or at a minimum of once per year.		

4. As Required Checks:

Check	Last Recorded	Current
Verify the cut-in and cut-out pressures		
Check operating temperature of the hydrocarbon scrubber		

Date	Comments, Checks, or Maintenance Performed:

Reviewed by: _____ Date: _____